

Alpha Resources, Inc. Certificate Of Analysis

AEB-2019

HEXACHLOROBENZENE OAS

LOT # 11949

Carbon and Chlorine

The values for Carbon was determined using a commercial CHN analyzer against Acetanilide SRM 141c, Batch No. 88080 from NIST.

Carbon

Carbon = 25.30%

The value for Chlorine was determined after oxygen flask combustion by argentometric titration against Chlorobenzoic Acid, SRM 2144 Batch No. 901708. This was obtained from NIST.

Chlorine

Chlorine = 74.70%

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

Prepared by D.M. Reasons
Director of Quality Assurance

