

Alpha Resources, Inc. Certificate Of Analysis

AR-1722

COAL STANDARD

LOT# 221202 LID# 761202

DRIED BASIS VALUE

MEAN WEIGHT % SULFUR	=	0.85
STANDARD DEVIATION	=	0.03
MEAN WEIGHT % ASH	=	21.50
STANDARD DEVIATION	=	0.35
MEAN WEIGHT % VOL. MATTER	=	21.80
STANDARD DEVIATION	=	0.80
BTU	=	12,140
STANDARD DEVIATION	=	65
FIXED CARBON	=	56.85

Methods Employed:

SULFUR	ASTM	D 4239 METHOD C
ASH	ASTM	D 3174
VOL. MATTER	ASTM	D 3175
BTU	ASTM	D 3286

Normal ASTM procedures should be employed when using this standard. This includes using the reproducibility and repeatability factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The material used in production of this standard was sampled in accordance with ARI 031. The above values relate only to the material used to produce this standard.

The overall direction and coordination of the analytical measurements leading to certification were performed by J.R. Shingledecker at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM). For good laboratory practice it is recommended that all standards be verified prior to use.

EXPIRATION DATE
THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF PURCHASE WHICH
APPEARS ON THE BOTTLE

P.O. Box 199 3090 Johnson Road Stevensville, MI 49127 USA Ph

www.alpharesources.com

Page 1 Of 1

J.R. Shingledecker

35-3629

Alpha Resources, Inc. Certificate Of Analysis

