

Alpha Resources, Inc. Certificate Of Analysis

AR-1906

COAL STANDARD

LOT # 906SUF LID ID SUF92

DRIED BASIS VALUE

WEIGHT PERCENT CARBON = 71.59
STANDARD DEVIATION = 0.29

WEIGHT PERCENT HYDROGEN = 4.76
STANDARD DEVIATION = 0.08

WEIGHT PERCENT NITROGEN = 1.17
STANDARD DEVIATION = 0.04

Methods

ASTM D 3178

ASTM D 3179

Notes:

Normal ASTM procedures should be employed when using this standard. This includes using the reproducibility and repeatability factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The statistical analysis and overall coordination leading to certification was performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

EXPIRATION DATE

THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF PURCHASE WHICH APPEARS ON THE BOTTLE



P.O. Box 199 3090 Johnson Road Stevensville, MI 49127 USA Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1

Alpha Resources, Inc. Certificate Of Analysis

