

Alpha Resources, Inc.

Certificate Of Analysis

AR 300
CAST IRON STANDARD
LOT # 080404

% CARBON
MEAN = 3.32
ONE SIGMA = 0.03
TWO SIGMA = 0.06
RANGE = 3.26 – 3.38

% SULFUR
MEAN = 0.014
ONE SIGMA = 0.001
TWO SIGMA = 0.002
RANGE = 0.012 – 0.016

Method of Analysis is ASTM E 1019-03

Primary Standards Employed:

NIST	4L, 338
JSS	111-12, 602-9
CTIF	FB-10-1
IPT	66

Notes:

The mean analytical values were derived by a number of data sets utilizing various ASTM approved instruments.

The precision values represent the standard deviation, two times the standard deviation, and complete range of analysis. When necessary, professional judgment is applied toward consideration of data and statistical information.

The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

The material used in production of this standard was sampled in accordance with ARI 032. The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Alpha Resources is an ISO/IEC 17025 accredited laboratory. For more information concerning our scope of accreditation contact Alpha Resources.

Certified August, 2004



Page 1 of 1

P.O. Box 199 3090 Johnson Road Stevensville, MI 49127 USA Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1