

Alpha Resources, Inc. Certificate Of Analysis

AR3153

LOT 153798

MEAN PPM NITROGEN = 10.01
STANDARD DEVIATION = 0.50

The certified mean value was statically derived from the values reported to Alpha Resources through Round Robin.

METHOD:
ASTM D 4629

NOTES:

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

This CRM is valid for two years from the date of OPENING.

The statistical analysis of the certification was performed by D.M.Reasons at Alpha Resources.

The overall direction and coordination of the analytical measurements leading to certification were performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a certified reference material (CRM) and should be verified, prior to use, against a primary standard material provided by a governing agency such as N.I.S.T, when available.

