

Alpha Resources, Inc. Certificate Of Analysis

AR-3704

COAL STANDARD

LOT # 704399LIG94 LID ID LIG94

ALL ANALYSIS REPORTED ON A DRIED BASIS VALUE

MEAN ug/g MERCURY	=	0.13
STANDARD DEVIATION	=	0.01
MEAN ug/g CHLORINE	=	107
STANDARD DEVIATION	=	50
MEAN WEIGHT PERCENT SULFUR	=	1.17
STANDARD DEVIATION	=	0.02
MEAN WEIGHT PERCENT ASH	=	10.31
STANDARD DEVIATION	=	0.23

Methods:

ASTM D 2031

ASTM D 3684-94

ASTM D 4208-97

ASTM D 4239 METHOD C-97

ASTM D 3174-97

Notes:

The certified mean values were statistically derived from the values reported to Alpha Resources through Round Robin.

The statistical analysis and overall coordination leading to certification was performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

EXPIRATION DATE

**THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF PURCHASE WHICH
APPEARS ON THE BOTTLE**



P.O. Box 199 3090 Johnson Road Stevensville, MI 49127 USA Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1