

Alpha Resources, Inc. Certificate Of Analysis

AR4000

MASONARY CEMENT STANDARD

LOT # 0517

MEAN WEIGHT % SULFUR = **2.64**
TRIOXIDE
STANDARD DEVIATION = **0.06**

Notes:

The certified mean values were statistically derived from the values reported to Alpha Resources through Round Robin.

Methods
ASTM D 5016
C 114

Standards Employed were:
NIST SRM 1886

The overall direction, coordination and statistical analysis were performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a certified reference material (CRM) and should be verified, prior to use, against a primary standard material provided by a governing agency such as N.I.S.T., when available.

