

Alpha Resources, Inc.

Certificate Of Analysis

AR-5201

JP-8 Jet Fuel

LOT # 5201

LID # 5201

ASTM	TEST	RESULTS
D 3242	Total Acid (Mg KOH/g)	0.0025
D 1319	Aromatics (vol.%)	17.70
D 1319	Olefins (vol.%)	2.45
D 2622	Total Sulfur. (Wt.%)	0.178
D 3227	Mercaptan sulfur (Wt.%)	0.0019
D 1298	API Gravity (60 ⁰ F)	41.85
D 2386	Freeze Pt. (⁰ F)	-42.75
D 3338	BTU (/lb.)	18536
D 1322	Smoke Pt. (mm)	21.0
D 93	Flash Pt. (⁰ F)	124.1
D 56	Flash Pt. (⁰ F)	117.4
D 445	Viscosity (-4 ⁰ F cSt)	6.14

The above value represents the mean from the data, as reported by Alpha Resources' round robin program.

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

Normal ASTM procedures should be employed when using this standard. This includes using the reproducibility and repeatability factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The material used in production of this standard was sampled in accordance with ARI 031. The above values relate only to the material used to produce this standard.

The overall direction and coordination of the analytical measurements leading to certification were performed by J.R. Shingledecker at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

EXPIRATION DATE
THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF PURCHASE WHICH
APPEARS ON THE BOTTLE

J.R. Shingledecker