Alpha Resources, Inc. Certificate Of Analysis

AR 555 HYDROGEN IN STEEL PIN STANDARD LOT # T39074C10

TOTAL HYDROGEN MEAN VALUE = 3.4 PPM (ug/g) ONE SIGMA DEVIATION = +/- 0.3 PPM (ug/g) TWO SIGMA DEVIATION = +/- 0.6 PPM (ug/g) TOTAL RANGE OF DATA = 2.8 to 4.0 PPM (ug/g)

> Method of Analysis: Inert Gas Fusion, Thermal Conductivity Detection

> > Primary Standards Employed: JSS SS-5-18

> > > Notes:

The mean analytical values were derived by a number of data sets, totaling n=40 utilizing the above mentioned method. The precision values represent the standard deviation, two times the standard deviation (k=2, 95% confidence limit), and the complete range of analysis data reported. When necessary, professional judgment is applied toward consideration of data and statistical information.

The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer, Technical Manager at Alpha Resources.

The material used in production of this standard was sampled in accordance with ARI 032. The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard. This bottle consists of 100, 1g pins, and has no expiration date.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Certified December 15, 2010

Vent er

Technical Manager

Page 1 of 1

P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269) 465-3629 www.alpharesources.com

Page 1 Of 1