

Alpha Resources, Inc.

Certificate Of Analysis

AR 660
OXYGEN & NITROGEN PIN STANDARD
LOT # T39074705

PPM OXYGEN
MEAN = 42
ONE SIGMA = 3
TWO SIGMA = 6
RANGE = 36 - 48

PPM NITROGEN
MEAN = 58
ONE SIGMA = 4
TWO SIGMA = 8
RANGE = 50 - 66

Method of Analysis is ASTM E 1019-03, ARI 034
Primary Standards Employed:

JSS	GS-1d, GS-3c, GS-5d
BAM	026-2
NCS	NS20012

Notes:

The mean analytical values were derived by a number of data sets utilizing instruments that meet the ASTM E1019 -03 requirements.

The precision values represent the standard deviation, two times the standard deviation, and complete range of analysis. When necessary, professional judgment is applied toward consideration of data and statistical information.

The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

The material used in production of this standard was sampled in accordance with ARI 032. The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard. This certificate cannot be reproduced except in full.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Alpha Resources is accredited to ISO/IEC 17025. For more information concerning our scope of accreditation contact Alpha Resources.

Certified August, 2005

