

Alpha Resources, Inc.

Certificate Of Analysis

AR 675
OXYGEN & NITROGEN PIN STANDARD
LOT # 052506

PPM OXYGEN
MEAN = 2160
ONE SIGMA = 200
TWO SIGMA = 400
RANGE = 1803 to 2519

PPM NITROGEN
MEAN = 3190
ONE SIGMA = 160
TWO SIGMA = 320
RANGE = 2853 to 3541

NOTE: The addition of graphite powder was used during analysis. The black/brown color of the pins is normal.

Method of Analysis is ASTM E 1019-03 and ARI 034
Primary Standards Employed:

JK	47
NIST	346a, 1090
INFM	Cu600

Notes:

The mean analytical values were derived by a number of data sets (n=50) by various instrumentation meeting ASTM E1019-03. The precision values represent the standard deviation, two times the standard deviation, and complete range of analysis. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

The material used in production of this standard was sampled in accordance with ARI 032. The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Alpha Resources is an ISO/IEC 17025 accredited laboratory. For more information concerning our scope of accreditation contact Alpha Resources.

Certified September, 2006



P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1