

Alpha Resources, Inc.

Certificate Of Analysis

COKE STANDARD

AR 723

LOT KOP97

DRIED BASIS VALUE

Mean Weight Percent Sulfur = 0.59
Standard Deviation = 0.03

The certified mean values were statistically derived from the values reported to Alpha Resources through Round Robin.

METHOD IS ASTM D 4239 METHOD C

NIST STANDARDS EMPLOYED WERE SRM 2683 AND 2684

COAL PREPARATION ASTM D 2013

This is a certified reference material (CRM) and should be verified, prior to use, against a primary standard material provided by a governing agency such as N.I.S.T, when available.

The statistical analysis and overall coordination leading to certification was performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This CRM is valid for two years from the date of purchase which appears on the bottle.

□

Alpha Resources, Inc. Certificate Of Analysis

