

Alpha Resources, Inc.

Certificate Of Analysis

AR 889

STEEL PIN STANDARD

LOT # 532997908

% CARBON
MEAN = 1.057
ONE SIGMA = 0.012
TWO SIGMA = 0.024
RANGE = 1.034 to 1.081

% SULFUR
MEAN = 0.0215
ONE SIGMA = 0.0005
TWO SIGMA = 0.0010
RANGE = 0.0205 to 0.0226

Method of Analysis is ASTM E 1019-03, and ARI 033

Primary Standards Employed:

NIST	16f, 163, 16e
BAM	227-1
JSS	602-9

Notes:

The mean analytical values were derived by a number of data sets (n=40). The precision values represent the standard deviation, two times the standard deviation (k=2, 95% confidence), and complete range of analysis. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

The material used in production of this standard was sampled in accordance with ARI 032. The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Alpha Resources is an ISO/IEC 17025 accredited laboratory. For more information concerning our scope of accreditation contact Alpha Resources.

Certified January, 2009

