

Alpha Resources, Inc. Certificate Of Analysis

AR1905

COAL STANDARD

LOT # 9051000 LID ID 905000

DRIED BASIS VALUE

WEIGHT PERCENT CARBON	=	77.498
STANDARD DEVIATION	=	0.506
WEIGHT PERCENT HYDROGEN	=	5.153
STANDARD DEVIATION	=	0.243
WEIGHT PERCENT NITROGEN	=	1.713
STANDARD DEVIATION	=	0.056

Methods

ASTM D 4239	ASTM D 3174
ASTM D 3286	ASTM D 3175

Notes:

Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The statistical analysis and overall coordination leading to certification was performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

EXPIRATION DATE

THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF PURCHASE WHICH APPEARS ON THE BOTTLE

