

Alpha Resources, Inc. Certificate Of Analysis

AR 309

NODULAR IRON STANDARD

LOT # 72301

% CARBON
MEAN = 3.75
ONE SIGMA = 0.03
TWO SIGMA = 0.06
RANGE = 3.69 – 3.81

% SULFUR
MEAN = 0.013
ONE SIGMA = 0.001
TWO SIGMA = 0.002
RANGE = 0.011 – 0.015

Method of Analysis is ASTM E 1019
(Latest Revision)

The certified mean values were derived by a number of data sets provided by various laboratories over different days by separate analysts using ASTM approved instruments.

The precision values represent the standard deviation, two times the standard deviation, and complete range of analyses.

Primary Standards Employed were:

NIST	SRM 41, 122f, 338
CTIF	FB10-1
JSS	111-12, 102-6
NCS	16011, 56005

Notes:


The statistical analysis of the certification was performed by K.E. Dyer at Alpha Resources.

The overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Certified August, 2001

P.O. Box 199 3090 Johnson Road Stevensville, MI 49127 USA  ax (269) 465-3629
www.alpharesources.cc
Page 1 Of 1