

Alpha Resources, Inc. Certificate Of Analysis

AR 659

OXYGEN & NITROGEN PIN STANDARD

LOT # STP95998

PPM OXYGEN
MEAN = 189
ONE SIGMA = 7
TWO SIGMA = 14
RANGE = 175 - 203

PPM NITROGEN
MEAN = 73
ONE SIGMA = 3
TWO SIGMA = 6
RANGE = 67 - 78

Method of analysis is ASTM E 1019
(Latest revision)

The following Primary Standards were used for calibration:

NIST	368, 1090
German	035-1
French	280-1

Notes:

The mean values reported were derived from a number of data sets utilizing ASTM approved instruments.

The precision values are based on the standard deviation, two times the standard deviation, and complete range of analyses.

The statistical analysis of the certification was performed by K.E. Dyer at Alpha Resources.

The overall direction and coordination of the analytical measurements leading to certification was performed by K.E. Dyer at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

Certified October, 1998

