

Alpha Resources, Inc.

Certificate Of Analysis

KED1022
CATALYST STANDARD
LOT # 1213B

% CARBON
MEAN = 24.0
One Sigma Standard Deviation = 1.3
Expanded Uncertainty = 2.6
(k=2, 95% confidence)

% SULFUR
MEAN = 12.6
One Sigma Standard Deviation = 0.6
Expanded Uncertainty = 1.2
(k=2, 95% confidence)

This data was reported using induction type furnaces with infrared detection. Accelerants like tungsten metal and iron chip were used in the induction analysis.

Standards Employed for traceability:

High Purity: CaCO₂, BaSO₄, Iron Sulfide
NCS DC70020
IRSID 701-1

Notes:

The mean analytical values were derived by 4 data sets (n=40) showing traceability to the above mentioned NMI and high purity standards, and reported in mass fraction. The precision values represent the estimated uncertainty derived from the data sets and may not represent your testing capabilities. Refer to your test method for the expanded method derived uncertainty if needed. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer, Technical Manager, at Alpha Resources.

There were limited primary standards of this type of matrix available at the time of certification. Multiple types of reference materials were used in the certification process for calibration and quality verifications. This bottle contains 100g of fine powder to be used directly from the bottle without preparation, and has an indefinite shelf life. Shake well before using.

The material used in production of this standard was sampled in accordance with ARI 032. The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Certified March 18, 2014



Technical Manager