

Alpha Resources, Inc.

Certificate Of Analysis

KEROSENE STANDARD

AR3117

LOT 117995 LID ID 3117

WEIGHT PERCENT SULFUR = LESS THAN 1 PPM

ROUND ROBIN SUMMARY

LAB	% SULFUR	METHOD	% REPORTED
1	< 1 PPM	1	.00009
2	< 1 PPM	1	.00003
3	< 1 PPM	1	.00004
4	< 1 PPM	1	.00001
5	< 1 PPM	1	.00000

METHOD
ASTM D 3246

NOTES:

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

The statistical analysis of the certification was performed by D.M. Reasons at Alpha Resources.

The overall direction and coordination of the analytical measurements leading to certification were performed by D.M. Reasons at Alpha Resources.

This is a secondary standard reference material, and should be verified against a primary standard such as NIST when possible.

EXPIRATION DATE

THIS CRM IS VALID FOR TWO YEARS FROM DATE OF OPENING



P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1