

Alpha Resources, Inc. Certificate Of Analysis

AR 148
COPPER PIN STANDARD
OXYGEN & SULFUR
LOT # 9163192205

PPM OXYGEN
MEAN = 489 PPM
ONE SIGMA = +/- 7 PPM
TWO SIGMA = +/- 14 PPM
RANGE = 477 to 500 PPM

PPM SULFUR
MEAN = 6 PPM
ONE SIGMA = +/- 1
TWO SIGMA = +/- 2
RANGE = 4 to 8 PPM

PPM CARBON = 13 PPM (REFERENCE ONLY)
PPM NITROGEN = 4 PPM (REFERENCE ONLY)

Method of analysis ARI-034 and ARI033

Oxygen	Inert Gas Fusion IR Detection
Sulfur	Combustion - IR Detection
Carbon	Combustion - IR Detection
Nitrogen	Inert Gas Fusion TC Detection

Standard Reference Materials employed:

NIST	396, 495, 496
BCS	260/4
INFM	Cu300, Cu600, Cu S20

Notes:

The mean analytical values were derived by a number of data sets (n=40) by various instrumentation meeting ASTM E1019-03. The precision values represent the standard deviation, two times the standard deviation, and complete range of analysis. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

The material used in production of this standard was sampled in accordance with ARI 032. The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Alpha Resources is an ISO/IEC 17025 accredited laboratory. For more information concerning our scope of accreditation contact Alpha Resources.

Certified August, 2006

