

# Alpha Resources, Inc.

## Certificate Of Analysis

AR 1652  
OXYGEN & NITROGEN PIN STANDARD  
LOT # 611C

**% OXYGEN**  
MEAN = 0.0045  
One Sigma Standard Deviation = +/- 0.0003  
Expanded Uncertainty = +/- 0.0007  
(k=2.3, 95% confidence)

**% NITROGEN**  
MEAN = 0.0530  
One Sigma Standard Deviation = +/- 0.0009  
Expanded Uncertainty = +/- 0.0018  
(k=2, 95% confidence)

Method of Analysis is ASTM E 1019-08, and ARI 034

Primary (NMI) Standards Employed:

NIST SRM	343a, 73c, 1093
BAM	O4, 026-2, 227-1
NCS	NS20035, NS22006
JSS	GS-3c, GS-1d

### Notes:

The mean analytical values were derived by 5 data sets (n=50) showing traceability to the above mentioned NMI standards, and reported in mass fraction. The precision values represent the estimated uncertainty derived from the data sets and may not represent your testing capabilities. Refer to your test method for the expanded method derived uncertainty if needed. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer, Technical Manager, at Alpha Resources.

The material used in production of this standard was sampled in accordance with ARI 032. The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard. This bottle contains 100g, 1.0g pins, to be used directly from the bottle with no preparation needed. This standard has an indefinite shelf life. Keep sealed and store in a cool dry place.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

This calibration standard is accredited and meets the requirements of ISO/IEC 17025 as verified by the ANSI-ASQ National Accreditation board. Alpha Resources is an ISO/IEC 17025 accredited laboratory. For more information concerning our scope of accreditation contact Alpha Resources.

Certified June 16, 2011



Technical Manager

P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269) 465-3629

[www.alpharesources.com](http://www.alpharesources.com)