

# Alpha Resources, Inc. Certificate Of Analysis

AR-1722

COAL STANDARD

LOT #722105 LID # 703105

DRIED BASIS VALUE

MEAN WEIGHT % SULFUR	=	0.84
STANDARD DEVIATION	=	0.03
MEAN WEIGHT % ASH	=	8.99
STANDARD DEVIATION	=	0.15
MEAN WEIGHT % VOL. MATTER	=	25.21
STANDARD DEVIATION	=	0.45
BTU	=	14073
STANDARD DEVIATION	=	55
FIXED CARBON	=	65.81

Methods Employed:

SULFUR	ASTM	D 4239 METHOD C
ASH	ASTM	D 3174
VOL. MATTER	ASTM	D 3175
BTU	ASTM	D 3286 & D 5865

Notes:

Sulfur, Ash, and BTU were traced to NIST SRM 2692b.

The material used in production of this standard was sampled in accordance with ARI 031. The above values relate only to the material used to produce this standard.

The overall direction and coordination of the analytical measurements leading to certification was performed by J.R. Shingledecker at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM). For good laboratory practice it is recommended that all standards be verified prior to use.

EXPIRATION DATE

**THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF PURCHASE WHICH  
APPEARS ON THE BOTTLE**

*J.R. Shingledecker*

Page 1 Of 1

P.O. Box 199 3090 Johnson Road Stevensville, MI 49127 USA Phone (269) 465-5559 Fax (269) 465-3629

[www.alpharesources.com](http://www.alpharesources.com)

Page 1 Of 1