

Alpha Resources, Inc. Certificate Of Analysis

AR-1908

COAL STANDARD

LOT # 908400 LID ID 908400

DRIED BASIS VALUE

WEIGHT PERCENT CARBON = 69.96
STANDARD DEVIATION = 0.41

WEIGHT PERCENT HYDROGEN = 4.83
STANDARD DEVIATION = 0.28

WEIGHT PERCENT NITROGEN = 0.92
STANDARD DEVIATION = 0.06

Methods
ASTM D 3178
ASTM D 3179

Notes:

Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The statistical analysis and overall coordination leading to certification was performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

EXPIRATION DATE

**THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF PURCHASE WHICH
APPEARS ON THE BOTTLE**

