

Alpha Resources, Inc. Certificate Of Analysis

AR1912

CHLORINE IN COAL

LOT #912297

LID ID 1912

DRIED BASIS VALUE

WEIGHT PERCENT CHLORINE = 0.3905
STANDARD DEVIATION = 0.0055

The certified mean values were statistically derived from the values reported to Alpha Resources through Round Robin.

Method
ASTM D 4208

Coal preparation ASTM D 2683a

Notes:

Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The statistical analysis and overall coordination leading to certification was performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

EXPIRATION DATE
THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF PURCHASE WHICH
APPEARS ON THE BOTTLE

