

Alpha Resources, Inc. Certificate Of Analysis

AR-1912

CHLORINE IN COAL

LOT #912802

LID ID 912802

DRIED BASIS VALUE

WEIGHT PERCENT CHLORINE = 0.24
STANDARD DEVIATION = 0.03

The certified mean values were statistically derived from the values reported to Alpha Resources through Round Robin.

Methods:

ASTM E 1151-93
ASTM D 4208

Coal preparation ASTM D 2683a

Notes:

Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The statistical analysis and overall coordination leading to certification was performed by J.R. Shingledecker

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

EXPIRATION DATE

**THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF PURCHASE WHICH
APPEARS ON THE BOTTLE**

J.R. Shingledecker

P.O. Box 199 3090 Johnson Road Stevensville, MI 49127 USA Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1