

Alpha Resources, Inc.

Certificate Of Analysis

CHLORINE IN PARAFFIN OIL STANDARD

AR2013

LOT 013894

MEAN WEIGHT PERCENT CHLORINE = 2.99
STANDARD DEVIATION = 0.04

The certified mean values were statistically derived from the values reported to Alpha Resources through Round Robin.

METHODS:

1. ASTM D 808-87 BOMB METHOD
2. ASTM D 4929

NOTES:

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

The statistical analysis of the certification was performed by D.M. Reasons at Alpha Resources.

The overall direction and coordination of the analytical measurements leading to certification were performed by D.M. Reasons at Alpha Resources

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a certified reference material (CRM) and should be verified, prior to use, against a primary standard material provided by a governing agency such as N.I.S.T, when available.

EXPIRATION DATE

THIS CRM IS VALID FOR TWO YEARS FROM DATE OF OPENING



P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com