

Alpha Resources, Inc.

Certificate Of Analysis

AR-2041

CRUDE OIL STANDARD

LOT # 411213

LID ID 411213

WEIGHT PERCENT SULFUR = 0.053

EXPANDED UNCERTAINTY = 0.005

(k=2, 95% confidence limit)

Method used for analysis: ASTM D 4294-10 and ARI-078

NMI Standard(s) used: NIST SRM 1623b, 1619a, 2717

This standard was produced gravimetrically using high purity materials, with balances calibrated and checked by precision NIST traceable weights. The above ASTM test method and NMI references listed were used for testing. The uncertainty value represents the 95% confidence limit (k=2) derived from analysis (n=30), and possible errors in weighing. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer, Technical Manager at Alpha Resources. Normal test procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors for the test method you wish to employ. The material used in production of this standard was sampled in accordance with ARI 032. The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard.

Notes:

Before use, the contents of the bottle should be mixed by vigorous shaking. Any exposure to air and light should be kept to a minimum. Keep sealed and store upright under normal laboratory conditions. This bottle contains 100ml to be used as per the test method. Unopened and stored properly this product has an indefinite shelf life. Once opened this standard is only valid for two years. This reference material was produced in accordance to ISO Guide 34-2009.


Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (working standard), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use. These test results are accredited under the Alpha Resources Inc. laboratory's ISO/IEC 17025 accreditation issued by ANSI-ASQ National Accreditation Board/ACCLASS. Refer to certificate and scope of accreditation AC-1200.

EXPIRATION DATE

THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING

Certified January 20, 2014



Technical Manager