

# Alpha Resources, Inc. Certificate Of Analysis

## PETROLEUM COKE STANDARD REFERENCE MATERIAL

AR2721

LOT 700 Lid ID 70092

% SULFUR = 6.15 .08

### METHODS

ASTM D 4239 METHOD C

ASTM D 4239 METHOD B

Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The statistical analysis and overall coordination leading to certification was performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

### EXPIRATION DATE

**THIS CRM IS VALID FOR TWO YEARS FROM DATE OF PURCHASE WHICH  
APPEARS ON THE BOTTLE**

