

Alpha Resources, Inc.

Certificate Of Analysis

AR-2783

ULTIMATE COAL STANDARD (ANTHRACITE)

LOT # 783410

LID # 783410

DRIED BASIS VALUES

Proximate Analysis		ASTM	Ultimate Analysis		ASTM
% Ash.....	17.58+/-0.14	D3174/D5142	% Carbon.....	77.07+/-0.7	D5373
% Volatile Matter.....	6.08+/-0.3	D3175/D5142	% Hydrogen.....	1.88 +/-0.12	D5373
% Fixed Carbon (calculated)	76.34	D3172	% Nitrogen.....	1.02 +/-0.26	D5373
% Sulfur.....	0.55+/-0.03	D4239B	% Oxygen (calculated)...	1.9	D5373
Btu.....	11990	D5865	MAF/DAF BTU	14548	D5865
%Mineral Analysis		ASTM	Sulfur Forms		ASTM
Silica.....	56.79 +/-0.6	D4326/D3682	% Pyritic.....	0.07	D2492
Alumina.....	29.58 +/-0.34	D4326/D3682	% Organic (calculated)...	0.43	D2492
Titanium Oxide.....	2.33 +/-0.18	D4326/D3682	% Sulfate.....	0.05	D2492
Ferric Oxide.....	5.20 +/-0.35	D4326/D3682	Ash Fusion Temperature		
Calcium Oxide.....	0.6 +/-0.1	D4326/D3682			
Magnesium Oxide.....	0.83 +/-0.07	D4326/D3682	ASTM D1857	Degrees F	Degrees F
Potassium Oxide.....	3.06 +/-0.1	D4326/D3682	Initial deformation...	Reducing	Oxidizing
Sodium Oxide.....	0.41 +/-0.04	D4326/D3682	Softening.....	<2640	<2730
Sulfur Trioxide.....	0.37 +/-0.09	D4326/D3682	Hemispherical.....	<2670	<2733
Phosphorus Pentoxide.....	0.16 +/-0.03	D4326/D3682	Fluid/Final.....	<2690	<2733
Strontium Oxide.....	0.03	D4326/D3682		<2700	<2733
Barium Oxide.....	0.14	D4326/D3682	% Chlorine.....	0.01	
Manganese Oxide...	0.03	D4326/D3682			
Undetermined (calculated)	0.47				

The material used in production of this standard was sampled in accordance with ARI 031. The uncertainty values represent the normal standard deviation ($k=1$, one sigma, 68% confidence) obtained through analytical testing, and may not reflect your testing capabilities. Normal ASTM procedures should be employed when using this standard, this includes using the *reproducibility* and *repeatability* factors of the method for establishing analytical uncertainty. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard. The analytical samples should be dried under a nitrogen atmosphere for a minimum of 70 minutes at 107° C +/- 3°C until a steady mass is achieved.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages. This is a Certified Reference Material (CRM). For good laboratory practice it is recommended that all standards be verified prior to use.

EXPIRATION DATE

THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING

CERTIFIED September 24, 2010



P.O. Box 199 3090 Johnson Road Stevensville, MI 49127 USA Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com