

Alpha Resources, Inc. Certificate Of Analysis

EXTREME PRESSURE OIL STANDARD

AR2852

LOT 522

% SULFUR = 3.30 +/- .06

ROUND ROBIN SUMMARY

LAB	% SULFUR	METHOD
1	3.24	1
2	3.33	1
3	3.30	1
4	3.36	1
5	3.27	1

METHODS:

1. ASTM 4294

STANDARD USED WAS NIST 1620a

NOTES:

THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF PURCHASE WHICH APPEARS ON THE BOTTLE.

BEFORE USE, THE CONTENTS OF THE BOTTLE SHOULD BE MIXED THROUGH VIGOROUS SHAKING. THE CONTENTS SHOULD NOT BE EXPOSED TO AIR FOR A LENGTHY PERIOD OF TIME.

THE STATISTICAL ANALYSIS OF THE CERTIFICATION WAS PERFORMED BY D.M. REASONS AT ALPHA RESOURCES.

THE OVERALL DIRECTION AND COORDINATION OF THE ANALYTICAL MEASUREMENTS LEADING TO CERTIFICATION WERE PERFORMED BY D.M. REASONS AT ALPHA RESOURCES.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.