

Alpha Resources, Inc.

Certificate Of Analysis

#1 DIESEL FLASHPOINT STANDARD

AR-3024

LOT# 024901 LID# 024901

FLASH POINT (D56) = 219⁰ FAHRENHEIT
STANDARD DEVIATION = 6⁰ FAHRENHEIT
FLASH POINT (D93) = 218⁰ FAHRENHEIT
STANDARD DEVIATION = 2⁰ FAHRENHEIT

ASTM METHODS:

ASTM D-56
ASTM D-93

The above value represents the mean from the data, as reported by Alpha Resources' round robin program.

NOTES:

Before use, the contents of the bottle should be mixed though vigorous shaking. This CRM should be refrigerated and tightly sealed when not in use. Any deviation may shorten shelf life.

Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The statistical analysis, and overall coordination leading to certification, was performed by D. M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

EXPIRATION DATE

THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF PURCHASE WHICH APPEARS ON THE BOTTLE

