

Alpha Resources, Inc. Certificate Of Analysis

KEROSENE VISCOSITY STANDARD

AR 3031
LOT 031893

VISCOSITY (in cSt at 40 degrees C) = 1.33 0.03

ROUND ROBIN RESULTS

| LAB | VISCOSITY | METHOD |
|-----|-----------|--------|
| 1 | 1.33 | 1 |
| 2 | 1.30 | 1 |
| 3 | 1.32 | 1 |
| 4 | 1.36 | 1 |
| 5 | 1.34 | 1 |

METHODS
1. ASTM 445

NOTES

Before using the contents of this bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

The statistical analyses and the overall direction and coordination leading to certification was performed by D.M. Reasons at Alpha Resources.

This is a secondary standard reference material, and should be verified against a primary standard such as NIST when possible.

EXPIRATION DATE
THIS SRM IS VALID FOR TWO YEARS FROM DATE OPENING



P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1