

Alpha Resources, Inc. Certificate Of Analysis

AR3121

GASOLINE STANDARD

LOT # 121995

LID ID 3121

WEIGHT PERCENT SULFUR = LESS THAN 1 PPM

Round robin summary

LAB	% SULFUR	METHOD	% REPORTED
1	< 1 PPM	1	0.00004
2	< 1 PPM	1	0.00003
3	< 1 PPM	1	0.00007
4	< 1 PPM	1	0.00001
5	< 1 PPM	1	0.00007

This Certified Reference Material was traced to N.I.S.T SRM 1623c

Notes:

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The statistical analysis of the certification was performed by D.M. Reasons at Alpha Resources.

The overall direction and coordination of the analytical measurements leading to certification were performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

EXPIRATION DATE

THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF PURCHASE WHICH APPEARS ON THE BOTTLE



P.O. Box 199 3090 Johnson Road Stevensville, MI 49127 USA Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1