

**AR 315**  
**CAST IRON SHOT STANDARD**

**LOT # G55**

**% CARBON**  
**MEAN = 3.09**  
**ONE SIGMA = 0.04**  
**TWO SIGMA = 0.08**  
**RANGE = 3.00 – 3.18**

**% SULFUR**  
**MEAN = 0.111**  
**ONE SIGMA = 0.005**  
**TWO SIGMA = 0.010**  
**RANGE = 0.100 – 0.121**

Method of Analysis is combustion / IR Detection ASTM E 1019-00  
(Latest Revision)

Primary Standards Employed were:

NIST SRM 107c, 4k, 122f  
CZECH CRM 215, 227  
Japan JSS 111-12  
France CRM FB-12

Notes:

The mean analytical values were statistically derived using a number of data sets provided by separate instruments over a number of days.

The precision values represent the standard deviation, two times the standard deviation, and complete range of analyses.

The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Certified January, 2001

