

Alpha Resources, Inc. Certificate Of Analysis

AR 4017

CARBON AND SULFUR IN SOIL

LOT # 40207

% CARBON
MEAN = 0.47
ONE SIGMA = 0.03
TWO SIGMA = 0.06
RANGE = 0.41 to 0.53

% SULFUR
MEAN = 0.52
ONE SIGMA = 0.03
TWO SIGMA = 0.06
RANGE = 0.46 to 0.58

Notes:

This data was reported using various induction and resistance type furnaces by IR detection. The use of accelerants like vanadium pent-oxide and tungsten tri-oxide were used in the resistance furnace techniques in order to properly liberate the carbon and sulfur. The use of tungsten and iron chip accelerators were used in the induction type analysis. Typical sample size was 200mg.


The data represents the mean value, one sigma standard deviation, two sigma standard deviation, and complete range of analysis. There were no primary standards of this type of matrix available at the time of certification. Many types of reference materials were used in the certification process, ranging from steel to high purity in-organics for calibration.

The statistical analysis and overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a reference material (RM) without any traceability to primary standards, and should be verified prior to use, against a primary standard material or other reference materials provided by a governing agency, such as N.I.S.T., when available.

Certified May, 2007



P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1