

Alpha Resources, Inc.

Certificate Of Analysis

AR 512
CARBON AND SULFUR PIN STANDARD
LOT # 511803107

% CARBON	% SULFUR
MEAN = 0.015	MEAN = 0.269
ONE SIGMA = 0.001	ONE SIGMA = 0.007
TWO SIGMA = 0.002	TWO SIGMA = 0.014
RANGE = 0.013 to 0.017	RANGE = 0.255 to 0.283

Method of Analysis is ASTM E 1019-03 and ARI 034

Primary Standards Employed:

NIST	SRM 2159, 133b, 129c, 368, 125b
BAM	284-2, 238-1, 297-1, 079-1
IPT	41B

Notes:

The mean analytical values were derived by a number of data sets (n=40) by various instrumentation meeting ASTM E1019-03. The precision values represent the standard deviation, two times the standard deviation, and complete range of analysis. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis, and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

The material used in production of this standard was sampled in accordance with ARI 032. The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Alpha Resources is an ISO/IEC 17025 accredited laboratory. For more information concerning our certificate and scope of accreditation contact Alpha Resources.

Certified May, 2007



Page 1 of 1