

Alpha Resources, Inc. Certificate Of Analysis

AR 546
HYDROGEN IN STEEL
LOT # 409303

TOTAL HYDROGEN
MEAN VALUE = 0.63 PPM (ug/g)
ONE SIGMA DEVIATION = +/- 0.2 PPM (ug/g)
TWO SIGMA DEVIATION = +/- 0.4 PPM (ug/g)
TOTAL RANGE OF DATA = 0.3 to .96 PPM (ug/g)

Method of Analysis:

1. LECO RH-404 Inert Gas Fusion
2. NRC Model 917 Vacuum Hot Extraction
3. LECO RH-402 Induction fusion

Standard reference materials used for certification:

JSS CRM SS-5-3, GS-1d, SS-5-18
NCS NS20025, NS20031

Hydrogen Gas Dosing

Notes:

Due to the extreme low hydrogen value, this standard is valued at the detection limit of most instrumentation, this is reflected in the large range of values and high standard deviation. The values represent a data set of n=50 numbers. Various analysts derived the mean analytical values using the above instruments over a number of days. The precision values represent the one-sigma standard deviation, two-sigma standard deviation and complete range of analytical data points used.

The material used in production of this standard was sampled in accordance with ARI 032. The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard. This certificate cannot be reproduced except in full. The statistical analysis, overall direction, and coordination of the analytical measurements leading to certification were performed by K. E. Dyer at Alpha Resources Inc.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Certified January, 2008



P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com