

Alpha Resources, Inc.

Certificate Of Analysis

AR 668
OXYGEN & NITROGEN PIN STANDARD
LOT # 409303609

PPM OXYGEN
MEAN = 14
ONE SIGMA = 1
TWO SIGMA = 2
RANGE = 12 to 16

PPM NITROGEN
MEAN = 31
ONE SIGMA = 2
TWO SIGMA = 4
RANGE = 27 to 35

Method of Analysis is ASTM E 1019-08, ARI 034

Primary Standards Employed:

JSS	GS-2c, GS-3c, 111-12, 4-80, 102-6
BAM	026-1, 079-1
NIST	1096, 1098
BCR	99

Notes:

The mean analytical values were derived by a number of data sets (n=40) following ASTM E1019-08 and ARI 034. The precision values represent the standard deviation, two times the standard deviation (k=2, 95% expanded uncertainty), and complete range of analysis. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

The material used in production of this standard was sampled in accordance with ARI 032. The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

This calibration standard is accredited and meets the requirements of ISO/IEC 17025 as verified by ACLASS the ANSI-ASQ National Accreditation board. Alpha Resources is an ISO/IEC 17025 accredited laboratory. For more information concerning our scope of accreditation contact Alpha Resources.

Certified May 10, 2010



Page 1 of 1