

Alpha Resources, Inc.

Certificate Of Analysis

AR 871

CARBON & SULFUR RING STANDARD

LOT # IW64745

% CARBON
MEAN = 0.053
ONE SIGMA = 0.001
TWO SIGMA = 0.002
RANGE = 0.051 – 0.055

% SULFUR
MEAN = 0.033
ONE SIGMA = 0.001
TWO SIGMA = 0.002
RANGE = 0.031 – 0.035

Method of Analysis is ASTM E 1019
(Latest Revision)

Primary Standards Employed:

NIST	SRM 345a, 11h, 14e
BCS	265/3
JSS	155-14

Notes:

The mean analytical values were derived by a number of data sets utilizing various ASTM approved instruments.

The precision values represent the standard deviation, two times the standard deviation, and complete range of analysis.

The statistical analysis of the certification was performed by K.E. Dyer at Alpha Resources.

The overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Certified November, 2001

