

Alpha Resources, Inc.

Certificate Of Analysis

AR 876
ALPHA RESOURCES RING STANDARD
LOT # T39074299

% CARBON
MEAN = 0.546
ONE SIGMA = 0.004
TWO SIGMA = 0.008
RANGE = 0.538 – 0.554

% SULFUR
MEAN = 0.0082
ONE SIGMA = 0.0006
TWO SIGMA = 0.0012
RANGE = 0.0070 – 0.0094

Method of Analysis is ASTM E 1019-94

Primary Standards Employed:

NIST	11h, 12h, 13g, 30b, 32b, 33a, 50, 50c, 73c, 106b, 152a, 161, 168, 367
BAS	306

Notes:

The mean analytical values were derived by a number of data sets utilizing various ASTM approved instruments.

The precision values represent the standard deviation, two times the standard deviation, and complete range of analysis. When necessary, professional judgment is applied toward consideration of data and statistical information.

The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

This standard was produced in accordance to Guide 34. Alpha Resources has become accredited under the ISO Guide 34:2009 for RMP and holds a ISO 17025 accreditation. Refer to certificate and scope of accreditation for details.

Certified November, 1999



Kent Dyer - Technical Manager