

# Alpha Resources, Inc.

## Certificate Of Analysis

AR 883  
STEEL PIN STANDARD  
LOT # A25909707

% CARBON  
MEAN = 0.190  
ONE SIGMA = 0.003  
TWO SIGMA = 0.006  
RANGE = 0.184 to 0.196

% SULFUR  
MEAN = 0.022  
ONE SIGMA = 0.001  
TWO SIGMA = 0.002  
RANGE = 0.020 to 0.023

Method of Analysis is ASTM E 1019-03 and ARI 033

Primary Standards Employed:

NIST	131c, 337a, 121, 12h, 73c, 367
JSS	030-7
NCS	NS 13016, 11011
BCS	238/2
IPT	26

### Notes:

The mean analytical values were derived by a number of data sets (n=60) by various instrumentation meeting ASTM E1019-03. The precision values represent the standard deviation, two times the standard deviation, and complete range of analysis. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer at Alpha Resources.

The material used in production of this standard was sampled in accordance with ARI 032. The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages. This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Alpha Resources is an ISO/IEC 17025 accredited laboratory. For more information concerning our scope of accreditation contact Alpha Resources.

This standard was produced in accordance to Guide 34 at the time of certification. These same methods for producing reference materials have now been reviewed by an accreditation body. As of February 2015 our facility has become accredited under the ISO Guide 34:2009 for RMP issued by ANSI-ASQ National Accreditation Board, certificate AR1920.

Certified November, 2007



Kent Dyer - Technical Manager

P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269) 465-3629

[www.alpharesources.com](http://www.alpharesources.com)