

Alpha Resources, Inc. Certificate Of Analysis

AR 959
STAINLESS STEEL CHIP STANDARD
LOT # 14833

% CARBON
MEAN = 0.019
ONE SIGMA = 0.001
TWO SIGMA = 0.002
RANGE = 0.017 – 0.021

% SULFUR
MEAN = 0.023
ONE SIGMA = 0.001
TWO SIGMA = 0.002
RANGE = 0.021 – 0.025

PPM NITROGEN
MEAN = 742
ONE SIGMA = 16
TWO SIGMA = 32
RANGE = 710 – 774

For reference only: Oxygen = 131 PPM

Method of Analysis is ASTM E 1019 (Latest Revision)

Standards employed:

NIST	SRM 367, 55b, 13g, 131c, 348, 125b, 343a, 8j
German	BAM 228-1, 238-1
Japanese	JSS 150-14, GS-1C
China	NCS 11301

Notes:

The mean analytical values were determined by a number of data sets provided by ASTM approved instruments.

The precision values represent the standard deviation, two times the standard deviation, and complete range of analyses.

The statistical analysis and overall coordination leading to certification was performed by K.E. Dyer at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a secondary standard reference material and should be verified, prior to use, against a primary standard material provided by a governing agency such as N.I.S.T., when available.

Certified June, 2000



P.O. Box 199 3090 Johnson Road Stevensville, MI 49127 USA Phone (269) 462-3629

www.alpharesources.com

Page 1 Of 1