Alpha Resources, Inc. Certificate Of Analysis

AR-2838

LUBE OIL STANDARD

LOT # 838808

LID # 838808

WEIGHT PERCENT SULFUR = 0.0100 STANDARD DEVIATION = 0.0004

The certified mean values were statistically derived from the values reported to Alpha Resources through Round Robin.

Methods: ASTM D 4294 ASTM D 5453

This Certified Reference Material was traced to N.I.S.T. SRM 1819 Oil I, II, III, IV, and V

Notes:

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The statistical analysis of the certification was performed by J.R. Shingledecker at Alpha Resources.

The overall direction and coordination of the analytical measurements leading to certification were performed by J.R. Shingledecker at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

The test results were validated on April 23, 2014 against NIST 1819a I, and II. These test results are accredited under the Alpha Resources Inc. laboratory's ISO/IEC 17025 accreditation issued by ANSI-ASQ National Accreditation Board/ACLASS. Refer to certificate and scope of accreditation AC-1200.

EXPIRATION DATE THIS CRM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING

J.R. Shingledecher