



## Certificate of Analysis

AR-1684

ULTRA LOW SULFUR COAL CRM

LOT # 684423

LID ID # 684423

### DRIED VALUE

MASS PERCENT SULFUR = 0.103

EXPANDED UNCERTAINTY =  $\pm 0.007$

(n=42)

METHOD: ARI-LAB-616

Combustion under Oxygen by Resistance Furnace with IR Detection \*ASTM D4239-18 (below method scope range)

This Reference Material is traced to:

NCS - FC28113, FC28003f, Alpha - AR1684-841019, AR1684-684414, AR1685-685414, AR1683-683610

The intended use of this certified reference material is for the calibration and quality validation of sulfur by resistance furnace combustion, infra-red detection analysis as specified by ASTM D4239 or other valid test methods. This Certified Reference Material (CRM) was prepared by gravimetric blending and verified by analysis. The typical sample size used and minimum sample size for analysis is approximately 300-500mg. The precision value represents the expanded degree of uncertainty based on errors from analytical assay at a 95% confidence level ( $k=2$ ) utilizing ISO Guide 35, ANOVA, and the Guide to Uncertainty Measurement. Metrological traceability is to the SI derived unit of mass fraction expressed as percent. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer, Chief Chemist at Alpha Resources. This CRM is below actual test method limits and no known NMI references at this concentration are available.

The material used for this standard was identified by ARI-LAB-603. The samples for testing were selected in accordance with ARI-LAB-625. The analytical samples were dried under a nitrogen atmosphere for a minimum of 30 minutes at  $107^{\circ}\text{C} \pm 3^{\circ}\text{C}$ , or until a steady mass is achieved. This bottle contains 50g fine -60 mesh (250 $\mu\text{m}$ ) powder to be used per your test method. While unable to determine a definite shelf life this CRM should be reviewed 20 years after the certification date. Once opened by the end user this certificate is valid for 2 years. Keep sealed immediately after use and store under normal laboratory conditions.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages. The above values relate only to the material used to produce this reference. This certificate cannot be reproduced except in full.

This is a CRM and is traceable to the above-mentioned reference material(s). For good laboratory practice, it is recommended that all reference materials be verified as fit for purpose prior to use. These test results are accredited under the Alpha Resources LLC laboratory's ISO/IEC 17025 and ISO 17034 (RMP) accreditation issued by ANSI-ANAB. Refer to certificate and scope of accreditation(s) AT-1200 and AR-1920.

### EXPIRATION DATE

THIS CRM SHOULD BE REVIEWED 20 YEARS FROM THE DATE OF CERTIFICATION

Certified April 13, 2023

Kent Dyer

Chief Chemist