

Alpha Resources, Inc.

Certificate Of Analysis

AR2004

CHLORINE IN PARAFFIN OIL STANDARD

LOT #004595

LID ID 004595

WEIGHT PERCENT CHLORINE = 0.0009
STANDARD DEVIATION = 0.0001

The certified mean values were statistically derived from the values reported to Alpha Resources through Round Robin.

Methods:
UOP-779
ASTM D 4929-89 METHOD B

Notes:

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The statistical analysis of the certification was performed by D.M. Reasons at Alpha Resources.

The overall direction and coordination of the analytical measurements leading to certification were performed by D.M. Reasons at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

EXPIRATION DATE
THIS CRM IS VALID FOR ONE YEAR FROM DATE OF OPENING



P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269) 465-3629

www.alpharesources.com

Page 1 Of 1