

Glycine Reference Material

Product No: AR211**Lot No: 241002**

Material and Intended Use

AR211 is an organic analytical reference material. The intended use of this RM is for the verification of combustion or other appropriate analysis methods for the determination of carbon, hydrogen, nitrogen, and oxygen. This RM can also be used to verify value assignment of in-house reference materials. A unit consists of one bottle containing 50 g of reference material as a fine powder. All reference materials should be verified as fit for purpose prior to use.

Instructions for Use

This product requires no preparation prior to use. The minimum sample size for analysis is dependent upon the test method and instrumentation used. It is recommended that no less than 0.25 g of material be used for destructive test methods. Bottles of powder should be kept sealed tight and stored in a cool, dry location. Reported values are valid until June 7, 2027 if handling and storage instructions are followed. The validity of property values is rendered null and void if the RM is in any way modified or damaged.

Reported Values

Assigned values indicate the amount of each element present in the overall material matrix and are metrologically traceable to the International System of Units (SI) derived unit of mass fraction expressed as a percent (%). Assigned values are reported as the mean property value derived from empirical stoichiometry and product purity (99.4%) consistent with guidance from the International Union of Pure and Applied Chemistry (IUPAC). Reported values are accredited under Alpha Resources, LLC ISO/IEC 17025 and ISO 17034 certificates issued by ANSI National Accreditation Board (ANAB), AT-1200 and AR1920.

Table 1. Reference values for AR211, Lot 241002.

Element	Value
%Carbon	31.81
%Hydrogen	6.67
%Nitrogen	18.55
%Oxygen	42.12

Homogeneity

This product was manufactured from a single lot of bulk material of purity 99.4%.

Methods and References

ISO/IEC 17025:2017 – General requirements for the competence of testing and calibration laboratories

ISO 17034:2016 – General requirements for the competence of reference material producers

ISO 33401:2024 – Reference materials – Contents of certificates, labels, and accompanying documentation

ISO Guide 30:2015 – Terms and definitions used in connection with reference materials

ISO Guide 35:2017 – Reference materials – General and statistical principles for certification

**Dustin Jenkins, Ph.D.****Global Technical Director****Certification Date:** October 3, 2024

This certificate cannot be reproduced except in full. Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources, LLC be liable for incidental or consequential damages. Produced in accordance with ISO 17025 and ISO 17034.