

Glycine Reference Material

Product No: AR211 Lot No: 250721

Material and Intended Use

AR211 is an organic analytical reference material. The intended use of this RM is for the verification of methods of elemental analysis for the determination of carbon, hydrogen, nitrogen, and oxygen. This RM can also be used to verify value assignment of in-house reference materials. A unit consists of one bottle containing 50 g of reference material as a fine powder. All reference materials should be verified as fit for purpose prior to use.

Instructions for Use

This product requires no preparation prior to use. The minimum sample size for analysis is dependent upon the test method and instrumentation used. It is recommended that no less than 0.25 g of material be used for destructive test methods. Bottles of powder should be kept sealed tight and stored in a cool, dry location. Reported values are valid until **June 23, 2028** if handling and storage instructions are followed. The validity of property values is rendered null and void if the RM is in any way modified or damaged.

Reported Values

Assigned values indicate the amount of each element present in the overall material matrix and are metrologically traceable to the International System of Units (SI) derived unit of mass fraction expressed as a percent (%). Assigned values are reported as the mean property value and an estimate of uncertainty derived from empirical stoichiometry and product purity (99.3%) consistent with guidance from the International Union of Pure and Applied Chemistry (IUPAC).

Table 1. Reference values for AR211, Lot 250721.

Property	Value	+/-
%Carbon	31.78	0.22
%Hydrogen	6.66	0.05
%Nitrogen	18.53	0.13
%Oxygen	42.33	0.30

Homogeneity

This product was manufactured from a single lot of bulk material of purity 99.3%.

Methods and References

ISO/IEC 17025:2017 – General requirements for the competence of testing and calibration laboratories
ISO 17034:2016 – General requirements for the competence of reference material producers
ISO 33401:2024 – Reference materials – Contents of certificates, labels, and accompanying documentation
ISO Guide 30:2015 – Terms and definitions used in connection with reference materials
ISO Guide 35:2017 – Reference materials – General and statistical principles for certification

Dustin Jenkins, Ph.D. Global Technical Director Certification Date: July 30, 2025



This certificate cannot be reproduced except in full. Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources, LLC be liable for incidental or consequential damages. Reported values are accredited under Alpha Resources, LLC ISO/IEC 17025 and ISO 17034 certificates issued by ANSI National Accreditation Board (ANAB), AT-1200 and AR1920.

3090 Johnson Rd. • Stevensville, MI 49127-0199 • Phone (269) 465-5559 • Fax (269) 465-3629 • alpharesources.com

Page 1 of 1 Rev. 2 - 2/19/2024