



Certificate of Analysis

AR-3085

ZINC SULFIDE CRM

LOT# N25G036

% SULFUR (BY PURITY)

32.9 %

Expanded Uncertainty = ± 0.003

Purity - 99.99%

% SULFUR (BY ANALYSIS)

MEAN = 32.9 %

Expanded Uncertainty = ± 1.1

($k=2.07$, 95% confidence) (n=23)

The intended use is for the Sulfur determination in materials using induction oxidation combustion with infrared detection or other valid test methods. The verification of this CRM was by induction oxidation combustion IR detection using tungsten metal and high purity iron chip for accelerators.

Reference material employed for traceability:

High Purity Zinc Sulfide 99.99% - Lot MKCJ1451

The mean analytical values were derived by data sets showing traceability to the above-mentioned reference and reported in mass fraction. Metrological traceability is to the SI derived unit of mass fraction expressed as percent. The precision values are derived using ISO Guide 35, the Guide to Uncertainty Measurement, and ANOVA. Refer to your test method or instrument manufacturer for the expanded method derived uncertainty if needed. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer, Chief Chemist, at Alpha Resources.

Sample size and minimum sample size for this data was 0.05g nominal. Refer to your instrument manufacturer or test method for minimum and typical sample size that can be used. This bottle contains 50g of fine powder to be used directly from the bottle without preparation. Keep sealed tightly and store under normal laboratory conditions. While unable to determine a definite shelf life this reference should be reviewed 15 years from certification.

The material used in production of this standard was sampled in accordance with ARI 032. The samples for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard. Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event, shall Alpha Resources be liable for incidental or consequential damages. This certificate cannot be reproduced except in its entirety.

This is a Certified Reference Material and is traceable to the above-mentioned standards. For good laboratory practice, it is recommended that all standards be verified as fit for purpose prior to use. This CRM was produced in accordance to ISO 17034 accreditation (RMP) issued by ANSI-ASQ/ANAB, refer to certificate AR-1920.

Certified May 29, 2020

Chief Chemist