

Alpha Resources, Inc. Certificate Of Analysis

NITROGEN IN VACUUM GAS OIL

AR3152

LOT 152697

MEAN PPM NITROGEN = 5.15

STANDARD DEVIATION = 0.66

The certified mean value was statically derived from the values reported to Alpha Resources through Round Robin.

METHOD:
ASTM D 4629

NOTES:

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

The overall direction and coordination of the analytical measurements leading to certification were performed by D.L. Andersson at Alpha Resources.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages.

This is a secondary standard reference material, and should be verified against a primary standard such as NIST when possible.

EXPIRATION DATE

THIS SRM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING



P.O Box 199 3090 Johnson Road Stevensville, MI 49127 Phone (269) 465-5559 Fax (269)

465-3629 www.alpharesources.com

Page 1 Of 1