



## Certificate of Analysis

KED1024

CARBON AND SULFUR IN CATALYST

LOT # 423P

**% CARBON**

MEAN = 15.2%

Standard Deviation =  $\pm 0.4$

Expanded Uncertainty =  $\pm 1.1$

(k=2, @ 95% confidence) (n=30)

**% SULFUR**

MEAN = 5.9%

Standard Deviation =  $\pm 0.4$

Expanded Uncertainty =  $\pm 0.9$

(k=2, @ 95% confidence) (n=30)

The intended use is for Carbon and Sulfur determination in alumina or silica-based catalysts or similar materials using induction combustion furnaces under oxygen with infrared detection. Tungsten metal and iron chip accelerators were used in the induction analysis. Method of analysis is similar to ASTM E1019-18, ASTM E1915-20 and ARI-LAB-621.

**Standards Employed for traceability:**

NCS

DC70010, DC70019, DC73508

ECISS

701-1

NIST

1d

ALPHA - KED1024-911C, KED1024-1018E, KED1022-911B

The mean analytical values were derived by data sets showing traceability to the above-mentioned NMI and Alpha standards and reported in mass fraction. The precision values are derived using ISO Guide 35, the Guide to Uncertainty Measurement, and ANOVA. Metrological traceability is to the SI derived unit of mass fraction expressed as percent. Refer to your test method or instrument manufacturer for the expanded method derived uncertainty if needed. When necessary, professional judgment is applied toward consideration of data and statistical information. The statistical analysis and the overall direction and coordination of the analytical measurements leading to certification were performed by K.E. Dyer, Chief Chemist, at Alpha Resources.

Analysis sample size and minimum sample size for this data was 100-150mg nominal. Refer to your instrument manufacturer for typical sample analysis size. This bottle contains 100g of fine powder to be used directly from the bottle without preparation. Keep sealed and store under normal laboratory conditions.

The material used in production of this standard was sampled in accordance with ARI-LAB-603. The samples for round robin testing were selected in accordance with ARI-LAB-625. The above values relate only to the material used to produce this standard.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Alpha Resources be liable for incidental or consequential damages. This certificate cannot be reproduced except in its entirety.

This is a Certified Reference Material (CRM) and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified as fit for purpose prior to use. These test results are accredited under the Alpha Resources LLC laboratory ISO/IEC 17025 and ISO 17034 accreditation (RMP) issued by ANSI-ASQ/ANAB. Refer to certificate and scope of accreditation(s) AT-1200 and AR-1920.

Reported values are valid for 10 years from the date of certification.

Certified July 26, 2023  
Dustin Jenkins, Ph.D.  
Global Technical Director